

## Certificate of Analysis

<b>Name of Product</b>	Tris Glycine Buffer (10X)
<b>Manufacturer</b>	Proteogenix, Pearls Business Park, Netaji Subhash Palace, New Delhi 110034, India
<b>Manufacturing Date</b>	23-04-2026
<b>Expiry Date</b>	23-04-2027

## Component

Description	Cat. No.	Lot No.
Tris Glycine Buffer (10X)	PX-TGB-601	PV6012412

**Performance Criteria** (Running or resolving Buffer for SDS-PAGE System)

**Specification:** Use as a running buffer for separating proteins in SDS-PAGE

## Quality Control Assessment

Test	Specification	Result
Colour	Transparent	Passed test
Appearance (Form)	Liquid	Passed test
pH	8.3	Passed test
Turbidity	Clear	Passed test
Suitability	Suitable for use as resolving buffer for SDS-PAGE	Passed test

**Storage Recommendations:** Store at Room Temperature

**Test Conducted by** Dr. Megha Kaushik **Testing Date** 30-04-2026

## Authorized Signature



Dr. Megha Kaushik  
Chief Scientific Officer  
Proteogenix Life science Private Limited

**Disclaimer:** This certificate certifies that the product meets the outlined specifications. We are committed to ensure that our products meet the highest quality standards, and we stand behind the efficacy and reliability of our offerings. If you have any questions or require further information, please reach out at technical services +919821999891 or from our website at [www.proteogenix.com](http://www.proteogenix.com). Actual results may differ based on experimental techniques and sample types. Users should perform their own validations tailored to their specific use case.

**Release Date:** 30-04-2026  
**Product Code:** PX-TGB-601

**Product Name:** Tris Glycine Buffer (10X)